

AD-A140 043


AN IMPROVED ION DRIFT METER(U) TEXAS UNIV AT DALLAS
RICHARDSON CENTER FOR SPACE SCIENCES B J HOLT NOV 83
UTD-23071-961(E-0618) AFGL-TR-83-0320 F19628-82-K-0041

1/1

UNCLASSIFIED


F/G 22/2

NL



100

[The remainder of the page is heavily obscured by a large black redaction mark.]



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

AFGL-TR-83-0320

AN IMPROVED ION DRIFT METER

B. J. Holt

Center for Space Sciences
The University of Texas at Dallas
Box 830688
Richardson, Texas 75083-0688

Final Report
24 May 1982 - 23 Nov 1983

November 1983

Approved for public release; distribution unlimited

DTIC FILE COPY

AIR FORCE GEOPHYSICS LABORATORY
AIR FORCE SYSTEMS COMMAND
UNITED STATES AIR FORCE
HANSCom AFB, MASSACHUSETTS 01731

DTIC
ELECTE
APR 11 1984
S D D

84 04 11 053

This report has been reviewed by the ESD Public Affairs Office (PA) and is releasable to the National Technical Information Services (NTIS).

"This technical report has been reviewed and is approved for publication"



WILLIAM P. SULLIVAN
Alt Contract Manager



WILLIAM J. BURKE Branch Chf
Space Plasmas & Fields Branch
Space Physics Division

FOR THE COMMANDER



RITA C. SAGALYN, Director
Space Physics Division

Qualified requestors may obtain additional copies from the Defense Technical Information Center. All others should apply to the National Technical Information Service.

If your address has changed, or if you wish to be removed from the mailing list, or if the addressee is no longer employed by your organization, please notify AFGL/DAA, Hanscom AFB, MA 01731. This will assist us in maintaining a current mailing list.

Do not return copies of this report unless contractual obligations or notices on a specific document requires that it be returned.

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER AFGL-TR-83-0320	2. GOVT ACCESSION NO. DD-A140043	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) An Improved Ion Drift Meter		5. TYPE OF REPORT & PERIOD COVERED Final Report 24 May 82 - 23 Nov 83
7. AUTHOR(s) B. J. Holt		6. PERFORMING ORG. REPORT NUMBER UTD 23071-961 (E-0618)
9. PERFORMING ORGANIZATION NAME AND ADDRESS Center for Space Sciences The University of Texas at Dallas P. O. Box 830688 Richardson, Texas 75083-0688		8. CONTRACT OR GRANT NUMBER(s) F 19628-82-K-0041
11. CONTROLLING OFFICE NAME AND ADDRESS Air Force Geophysics Laboratory Hanscom AFB, Mass. 01731 Monitor: Frederick J. Rich/PHG		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS 62101F 760112AQ
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office) Office of Naval Research 300 East 8th Street Austin, Texas 78701		12. REPORT DATE November 1983
		13. NUMBER OF PAGES 49
		15. SECURITY CLASS. (of this report) Unclassified
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Space Plasma, Ion Drift, Retarding Potential Analyzer		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) This report provides the necessary operation, diagnostic and repair information for using and maintaining the Improved Ion Drift Meter (IDM). The report is divided into four sections. Section 1 contains the instrument description and theory of operation. Section 2 contains a description of planned spacecraft-level instrument testing, stimulation requirements and procedure, and instrument handling and safety. Section 3 contains a copy of the End-Item Acceptance Test (EIAT) that describes test results and presents calibration data and procedures. Any recalibration of the instrument will be performed per the EIAT calibration test [over]		

DD FORM 1 JAN 73 1473

EDITION OF 1 NOV 65 IS OBSOLETE

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

procedure. Section 4 contains a list of the schematics, drawings, parts lists and wiring lists that describe the as-built configuration of the instrument. This documentation is available to AFGL upon request.

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

Preface

↙ This report provides the necessary operating, diagnostic and repair information for using and maintaining the Improved Ion Drift Meter (IDM). The report is divided into four sections. Section 1 contains the instrument description and theory of operation. Section 2 contains a description of planned spacecraft-level instrument testing, stimulation requirements and procedure, and instrument handling and safety. Section 3 contains a copy of the End-Item Acceptance Test (EIAT) that describes test results and presents calibration data and procedures. Any recalibration of the instrument will be performed per the EIAT calibration test procedure. Section 4 contains a list of the schematics, drawings, parts lists and wiring lists that describe the as-built configuration of the instrument. This documentation is available to AFGL upon request. ↘

Accession For	
NTIS GRA&I	<input checked="checked" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By _____	
Distribution/	
Availability Codes	
Dist	Avail and/or Special
A/1	



Table of Contents

	<u>Page</u>
INSTRUMENT DESCRIPTION	1-1
GENERAL DESCRIPTION	1-1
THEORY OF OPERATION	1-1
INSTRUMENT DETAILS	1-8
Telemetered Data	1-8
Data Storage/Readout	1-8
Power	1-11
Timing Signals from S/C to IDM	1-11
Analog Data Lines from RPA to IDM	1-12
Timing Signals from IDM to RPA	1-12
S/C Interface Circuits	1-12
IDM Data Interpretation	1-15
SPACECRAFT-LEVEL INSTRUMENT TESTING	2-1
Electrical Performance Evaluation Test	2-1
Stimulation	2-1
Sensor Grid Test	2-1
INSTRUMENT HANDLING AND SAFETY	2-2
RED-TAG ITEMS	2-2
STIMULATION PROCEDURE	2-3
INSTRUMENT CHECKOUT PROCEDURE	2-4
CALIBRATION DATA AND PROCEDURES	3-1
IDM DRAWING LIST	4-1

Referenced Illustrations

	<u>Page</u>
IDM Sensor Cross-Section	1-2
IDM Functional Diagram	1-3
Drift Meter Bias Diagram	1-7
Data Sampling Format	1-9
Drift Meter Timing	1-10, 1-13
Drift Meter/SDF Interface	1-14
HILAT IDM Collector Configuration	3-7
DSM Drawing List	4-1

Contributing Scientists/Engineers

The following scientists and engineers contributed to the research reported in this document:

W. B. Hanson
R. A. Heelis
C. R. Lippincott
D. R. Zuccaro
B. J. Holt
L. L. Harmon

Address:

The University of Texas at Dallas
Mail Stop F022
P. O. Box 830688
Richardson, Texas 75083

Telephone:

214-690-2851

or

214-690-2832

SECTION 1
INSTRUMENT DESCRIPTION


1.0 GENERAL

The IDM is designed to be combined with a retarding potential analyzer (RPA) built by the Air Force Geophysics Laboratory (AFGL). The combined IDM/RPA is part of the instrument complement designed for the DNA/TRANSIT HILAT satellite. The IDM provides all telemetry and timing signal interface to the spacecraft, timing signals to RPA, data storage and read-out. Since the positive potentials exposed on the spacecraft solar array are expected to drive the spacecraft potential negative, the IDM provides a scheme for electrically isolating the sensors and actively maintaining the sensor potential near the plasma potential.

2.0 THEORY OF OPERATION

The drift meter sensor is illustrated in Figure 1. The sensor consists of a square aperture placed in front of a collector that is segmented into quadrants. The first four grids are fixed at the floating ground potential to provide a field free drift space for the ions. The last grid is biased negatively and serves as an electron suppressor for the collector and also as an electron repeller for ambient electrons.

The ion collector consists of four pie-shaped segments fitted together into a circle. Each 90° sector is insulated from the others and from the sensor housing. For reference purposes, the collector plates are labeled A, B, C, and D (Figure 2). Pairs of collector plates are selectively coupled to the electrometers (log-amps) by means of the four switches labeled S1 through S4. (Actually, all switches in this diagram are FET transistors with appropriate gate bias voltages.) The log-amps are labeled + (A2) and - (A3) for reference purposes. For the switch settings shown, plates A and D are connected to the + log-amp, while plates B and C are connected to the - log-amp. When




GRID DESIGNATORS

- G1- INPUT (DUAL)
- G2- SHIELD
- G3- SHIELD
- G4- SUPPRESSOR

IDM SENSOR CROSS-SECTION

FIGURE 1



DRIFT METER FUNCTIONAL DIAGRAM

FIGURE 2

the spacecraft is properly oriented, and the ions are drifting upward go through the aperture, plates A and D receive more ions than plates B and C. This causes more positive current to flow into the + log-amp than into the - log-amp. The outputs of both A2 and A3 are negative but A2 is more negative because of the extra current. (A2 and A3 are inverting feedback amplifiers). The 2N2060 dual NPN transistor between A2 and A3 is a logarithmic feedback device; therefore the output voltages of A2 and A3 are proportional to the logarithm of the input currents.

The outputs of the log-amps are compared in the difference amplifier A5. The output of A5 then, depends only on the ratio of input currents to the log amps, and the polarity of the A5 output corresponds to the direction of the ion drift.

The 2N2060 transistor used as a logarithmic element contains two matched transistors in one case. Since the units are matched and are at the same temperature, the non-logarithmic constant and temperature effects will be cancelled over the input current range and temperature range. There remains a temperature effect on logarithmic sensitivity. This is easily compensated by resistors in the circuitry associated with A5. The error from these effects is well less than one percent over the total range of the instrument.

The transient muting circuit (S6) is needed to eliminate large transients from the signal path of the instrument during the time when the collector plates are being switched from one input to the other. They die out rather rapidly (milliseconds) but their presence at the filter could cause appreciable error in the resulting output without this muting circuit.

The log-level circuit at the top of the figure monitors the ion current level at the input collector plates. Switch S5 selects either the + or -

or the - log-amp and compares this to a reference voltage developed by A1. The reference log amp (A1) has a fixed input current. Difference amplifier A4 subtracts the output of the lower log amp from the reference voltage and amplifies the difference to produce an output corresponding to the amplitude of the ion input current. S5 is switched such that the + and - log amps are alternately sampled. This data is used primarily to monitor the behavior of the individual log-amps over the mission lifetime but also provides a measurement of N_i .

The drift signal, which is proportional to the tranverse components of the ion convection velocity, is developed by first filtering the output of A5 to limit the band width to prevent aliasing at the telemetry sample rate. A three pole low-pass Butterworth filter is used. The signal is then capacitively coupled into amplifiers A7 and A8. The capacitive coupling allows amplifier rezeroing to compensate for imbalances in the amplifiers and for amplifier drift due to temperature and aging.


Amplifier A8 and the associated input capacitor are used for horizontal (H) measurements while A7 and its capacitor are used for vertical (V) measurements.

At the beginning of a measurement sequence the collectors are configured to, say, the V configuration, the V zero switch (S7) is closed and the output of A7 is connected to the ranging circuit by S9. The resistor network at the output of A7 established the drift signal zero level. The zero level is nominally +2.5 V which corresponds to the center of the telemetry band. The V coupling capacitor is charged to a voltage which represents the drift angle during the re-zero time. After the V zero measurement is made, the zero switch is opened and a measurement is made with the collector plates reversed. The voltage out of the filter goes to the opposite polarity but with the same amplitude. Since the V capacitor has its original charge that was established

during the re-zero measurement, and the new voltage from the filter is equal but opposite, the voltage into amplifier A7 is twice the voltage corresponding to the drift angle. This is referred to as the V offset or double angle measurement. The automatic ranging circuit adjusts the gain of the amplifier for maximum drift signal amplification while maintaining the output within the telemetry band.

After the V offset measurement is made, the collectors are set to the horizontal configuration and H rezero and offset measurements are made using A8 and the H capacitor to preclude disturbing the charge on the V capacitor. Both capacitors are now charged to voltages proportional to the V and H ion angle of arrival at the times that the offset measurements were made. The capacitor and amplifier circuits are designed such that the capacitor charges can be maintained for several seconds, so by setting the collectors to either the H or V rezero configuration (without closing S7 or S8), and connecting S9 to the appropriate amplifier (A7 or A8), small changes in ion arrival angle can be observed relative to the "zeroed" angle on the most sensitive amplifier range. Either V or H arrival angle changes can be observed for several seconds at the available TM sample rate or alternate V and H measurements can be made using S9 to switch between the A7 and A8 outputs.

The scheme used for maintaining the sensor and associated electronics near the plasma potential is illustrated in Figure 3. The ground plane, sensor housings, electronics box and outer bundle shield of the sensor-to-electronics box cable are isolated from the spacecraft chassis. The "floating ground" potential on these elements of the instrument is established in-flight when equal ion and electron currents are being collected by the forward-looking sensor faces and ground plane. (This equilibrium should occur at ≈ 1 volt



NOTE:
ELECTRONICS BOX AND SENSOR MOUNT
ARE IN ELECTRICAL CONTACT WITH S/C
CHASSIS. SENSORS ARE INSULATED FROM
THE MOUNT AND GROUND PLANE IS
INSULATED FROM SENSORS.

DNA / SAT DRIFT METER BIAS DIAGRAM

FIGURE 3

below the plasma potential.) The instrument is powered by floating power supplies referenced to the floating ground potential. Level shifters are used for interfacing digital signals with the spacecraft and DC isolation of the power system is obtained through the DC to DC converter transformer. A low-leakage zener diode is used between floating ground and spacecraft ground to preclude large floating ground potentials during ground operations.

3.0 INSTRUMENT DETAILS

3.1 Telemetered Data

- RPA electrometer (RPA)
- RPA sweep monitor (SWP)
- Ion Drift Meter angle of arrival measurements (IDM)
- Frame Counter (COUNT) - counts TM frames in modulus 128 (0 to 127)
- IDM log electrometer outputs (LLA and LLB)
- Electron probe electrometer (ELECT)
- Temperature monitor (TEMP)
- Sensor potential monitor (SENPOT)
- 4 bandpass filter outputs (F1, F2, F3, F4)

3.2 Data Storage/Readout

• The data format is given in Figure 4. Each telemetry frame is divided into 64 equal time slots by a 128.0625 hz word counter. The telemetry words are numbered according to the time slot in which they occur. The data is sampled and stored 5.06 msec after the beginning of the time slot as indicated in Figure 5. The 512 bits/frame are allocated as follows:

FIGURE 4

COUNT	DNA/SAT SUBCOM	DATA ELECT	SAMPLING IDM	FORMAT PPA	(Nov. 3, 1982) SWP
0	0				
1		1			
2			2		
3				3	
4					
5		5			
6			6		
7				7	
8					8
9		9			
10			10		
11				11	
12					
13		13			
14			14		
15				15	
16	16				
17		17			
18			18		
19				19	
20					
21		21			
22			22		
23				23	
24					24
25		25			
26			26		
27				27	
28					
29		29			
30			30		
31				31	
32	32				
33		33			
34			34		
35				35	
36					
37		37			
38			38		
39				39	
40					40
41		41			
42			42		
43				43	
44					
45		45			
46			46		
47				47	
48	48				
49		49			
50			50		
51				51	
52					
53		53			
54			54		
55				55	
56					56
57		57			
58			58		
59				59	
60					
61		61			
62			62		
63				63	




FIGURE 5

<u>Word Content</u>	<u># Words/frame</u>	<u># Bits/word</u>	<u>Total # bits</u>
RPA	32	9	288
SWP	4	8	32
IDM	16	10	160
SUBCOM	<u>4</u>	<u>8</u>	<u>32</u>
TOTALS	56		512

- The data content of the subcom words is multiplexed as follows:

<u>Word #</u>	<u>Frame #</u> <u>0,4,8 ...124</u>	<u>Frame #</u> <u>1,3,5...127</u>	<u>Frame #</u> <u>2,6,10...126</u>
0	COUNT	F1	COUNT
16	LLA	F2	LLA
32	LLB	F3	LLB
48	TEMP	F4	SENPOT

• Two data storage registers are used so that one register may be loaded while the other is read by the telemeter. Data readout will lag data store by 0.49976 secs. (1 frame). Each word is serially stored as it occurs in the data format. Since the telemeter is an 8 bit system, the 8,9 and 10 bit word patterns will have to be reconstructed on the ground.

3.3 Power

The UTD-built portion of the instrument (IDM) requires the following voltages (F = floating):

+ 12 VF	@ 35 ma
- 12 VF	@ 45 ma
+ 5 VF	@ 15 ma
+ 5 V	@ 10 ma

3.4 Timing Signals from S/C to IDM

- 4098 Hz clock (CLK)
- Read Out Gate (ROG)
- SOF - This signal probably will not be used.

3.5 Analog Data Lines from RPA to IDM


- RPA - A single line connected to the RPA electrometer during frames 0-123 and to the electron probe electrometer during frames 124-127. Multiplexing is done within the RPA electronics.
- SWP - A single line connected to the RPA sweep monitor continuously.
- ELECT - A single line connected continuously to the electron probe electrometer.
- TEMP - A single line connected continuously to the RPA temperature monitor.
- FILTERS - A single line that is sequentially connected to the filter outputs such that filters F1, F2, F3 and F4 are sequentially sampled by words 0, 16, 32 and 48. This multiplexing is done in the RPA electronics. Additional multiplexing in the IDM electronics will allow the filter data to be sampled only during odd numbered frames.

3.6 Timing Signals from IDM to RPA (Logic level 1 = + 5 volts)

- RPA CONVERT - (See Figure 3) A line with ≈ 32 Hz pulses of ≈ 25 μ sec duration. The pulses are high during each A/D conversion of the RPA line.
- RPA SYNC - A line with a single pulse of ≈ 60 μ sec duration during frame 0 only.
- POR - A power on reset pulse that goes high at power on and goes low at the next ROG. An RPA SYNC pulse is always generated after the POR pulse as shown in Figure 6.



3.7 S/C Interface Circuits

Interface circuit sketches are attached. (Figure 7)



TIMING DIAGRAM

FIGURE 6



DRIFT METER/SDF INTERFACE

FIGURE 7

3.8 IDM Data Interpretation

• There are two drift signal amplifier channels, each channel having 4 sensitivity ranges differing by a factor of 4. At the beginning of each 4 second period, H and V rezero (H_z, V_z) and offset (H_o, V_o) measurements are made with the repeller grid bias set at $\sim +2$ volts. The grid is then set to zero and the H and V rezero and offset measurements are repeated. The H and V rezero values are stored independently and deviations (H,V) from these values are telemetered in the following manner:

<u>Beginning at</u>	<u>BIAS ON</u>	<u>BIAS OFF</u>	
Frame 0	$H_z H_o V_z V_o$	$H_z H_o V_z V_o$	H H H H ...
Frame 8	$H_z H_o V_z V_o$	$H_z H_o V_z V_o$	H V H V ...
Frame 16	$H_z H_o V_z V_o$	$H_z H_o V_z V_o$	H H H H ...
Frame 24	$H_z H_o V_z V_o$	$H_z H_o V_z V_o$	H V H V ...

• Each sample is 10 bits, 2 bits for range data and an 8 bit A/D conversion of the drift signal.

• The LLA and LLB TM Samples always occur in the horizontal rezero configuration, i.e. collectors A & B are connected to LLB and collectors C & D are connected to LLA.

SECTION 2

SPACECRAFT-LEVEL INSTRUMENT TESTING,

INSTRUMENT HANDLING AND SAFETY

1.0 General

This section defines the electrical tests planned for the Driftmeter/RPA after installation on the spacecraft.

2.0 Electrical Performance Evaluation Test (EPET)

The EPET provides an automated test procedure for accumulating up to 64 consecutive seconds of instrument data and printing the data in a specified format. The instrument performance evaluation will be based on visual observation of the printed data. The EPET can be performed either with or without stimulation. The detailed procedure for performing the EPET utilizing the HP9825 computer is given in 8.0 below.

3.0 Stimulation

Battery powered stimulation units are used to provide test currents to the sensors. The test currents are injected by probes that screw into the front face of the sensor and make contact with the collectors. The stimulation unit requires ~6 inches clearance in front of the sensor aperture plane for installation. The planned usage of the stimulation unit is limited to tests performed with attending UTD/AFGL personnel. Installation or removal of the stimulation unit will require approximately 30 minutes.

4.0 Sensor Grid Test

The sensor grid test is performed by UTD/AFGL personnel after final instrument installation on the spacecraft, after completion of S/C environmental testing, and during S/C checkout at WTR. The test consists of attaching test probes to the internal sensor grids and verifying the correct grid potentials.

5.0 Instrument Handling and Safety

5.1 Protective sensor covers should be utilized whenever sensors are installed (except during EMI and vibration testing) to prevent particulate contamination from entering the sensor entrance apertures and to prevent physical damage to the exposed entrance grids. Stimulus probes may be used with the protective covers in place.

5.2 Always minimize instrument handling and utilize approved procedures when handling is necessary.

6.0 Red Tag Items

 The protective covers over the sensor apertures will be removed at the latest practical time at WTR. Captive hardware is used for securing the protective covers to the sensor face. After removal of the protective covers, gold plated screws must be installed in the sensor face to fill the holes.


7.0 STIMULATION PROCEDURE

7.1 If sensor assy is not connected to the E-box, connect stimulation unit to E-box as follows:

- Stimulation connector "C" to J3 (SENPOT) connector
- Stimulation connector "-" to J4 - A1 (collector B)
- Stimulation unit connector "+" to J4 - A3 (collector D)

7.2 If sensor assy is connected to E-box, connect stimulation unit as follows:

- Install stimulation probes in sensor face to collectors B and D.



- Connect stimulation unit connector "C" center conductor to the sensor ground plane and connect "C" shield to S/C chassis.
- Connect stimulation unit connector "+" to collector D and "-" to collector B.

7.3 Set "FIX/CYCLE" switch on stimulation unit to "FIX".

7.4 Set "ON/OFF" switch to "ON"

7.5 Record and print 16 frames of IDM data.

7.6 Compare data to sample print-out. In particular, note the following

(1) LLA \cong 3.48

LLB \cong 3.22

(2) H offset = 1.68/4

V offset = 3.40/4

H rezero = 2.54/1

V rezero = 2.54/1


(3) SENPOT LOW = 4.52

SENPOT HIGH = 2.38

[battery voltage = 18-4 (SENPOT voltage)]

8.0 Instrument Checkout Procedure


- Step 1: Connect power cords on each unit to 110 V A.C. (9876A, 9825, AGE)
- Step 2: Connect 98032A interface cable between 9825 computer and AGE.
(top connector on back of AGE unit)
- Step 3: Connect 98034A interface cable between 9825 computer and 9876A printer.
- Step 4: Connect braided interface cable between AGE and DNA instrument.
- Step 5: Turn on power to each unit. (switch locations: left front on 9876A, right side on 9825, center front on AGE and left front on AGE)
Set both thumbwheel switches to "15,FF".
- Step 6: Load magnetic tape cartridge (RAP009) into 9825 tape drive.




Insert the tape cartridge so the label on the cartridge faces the back of the calculator as shown.

Inserting the Tape Cartridge

- Step 7: Press "LOAD" system command key. Press "EXECUTE" key. Checkout program will then load into computer memory.



Step 8: Press "RUN" key to start program. Flag definitions will be printed on twenty-column printer.



Step 9: Change flag settings, if desired, by using "sfg" and/or "cfg" commands. For example: to set flags 0 and 3, type

sfg 0,3

then press the "EXECUTE" key.

To clear flags 1 and 2, type

cfg 1,2

then press the "EXECUTE" key.


See Appendix A for flag definitions.

Step 10: Enter date, time, and operational comments by typing in as desired from keyboard followed by "CONTINUE". Each line of input is printed on the 9876A printer when the "CONTINUE" key is pressed. As many lines of comment may be entered as desired. Terminate the input session by again pressing "CONTINUE" without entering any further comments. (If a mistake is made press "CLEAR" key to clear the input.)

Step 11: Enter starting major frame no. and ending major frame no. to be printed. Separate the frame numbers with a comma. Press "CONTINUE" key to proceed. (If a mistake is made press "CLEAR" key to clear the input.)

Data transfer will then be initiated by program, followed by data print-out on 9876A.

Special Function Keys



Alphanumeric Keys

Numeric Keys

Step 12: Program cycles back to step 11 for next test if desired. To terminate session at any time press "RESET" system command key.

Step 13: To re-start program after a termination (see step #12), proceed from step #8 in instructions.

Step 14: Remove magnetic tape cartridge by pressing eject button directly below cartridge.

Step 15: Turn off power to each unit.

Appendix A

flag no.	definition	default
0	record data on magnetic tape	off
1	print RPA data	on
2	print IDM data	on
3	off = major frame no. input in step #11 refers to the data transfer frame no., i.e. frame no. 1 is the first frame recorded by the program on = major frame no. input in step #11 refers to the data frame count generated by the instrument	off
9	program test pattern generator	off

SECTION 3

CALIBRATION DATA AND PROCEDURES

REVISIONS			
LTB	DESCRIPTION	DATE	APPROVAL

AUTHENTICATION		THE UNIVERSITY OF TEXAS AT DALLAS			
NAME	DATE				
DRN.		HILATION DRIFT METER END-ITEM ACCEPTANCE TEST			
CHE.					
DESIGN <i>11-42-17</i>	<i>11-7-12</i>				
APP.					
APP.					
APP.					
3-1		A	CODE IDENT 14614	133-004	REV
		SCALE	WT	SHEET 1	OF 10

1.0 SCOPE

This document establishes the acceptance performance test criteria for end-item testing (EIAT) at UTD of the ION DRIFT METER (IDM) for the DNA HILAT satellite.

2.0 APPLICABLE DOCUMENTS

UTD 131-012 Test Requirements, General
IDM Schematics, Current Shop Copy
UTD 133-003 HILAT IDM Interface Document

3.0 GENERAL REQUIREMENTS

3.1 PERSONNEL

The GSE shall be operated and the testing performed by the IDM Electronics Engineer and/or qualified personnel designated by him.

3.2 TEST ENVIRONMENT

The Electrical Performance Evaluation Test (EPET) shall be performed at $+60\pm4^{\circ}\text{C}$, $+25\pm4^{\circ}\text{C}$ (RT), and $-20\pm4^{\circ}\text{C}$ at room pressure. The Calibration Test shall be performed at $+50\pm4^{\circ}\text{C}$, $+25\pm4^{\circ}\text{C}$, and $0\pm4^{\circ}\text{C}$.

The vacuum source tests shall be performed at room temperature and at pressures less than 2×10^{-5} torr (nominal) as measured on the ionization gauge. There is no requirement for calibrating the ionization gauge.

3.3 DATE TEST PERFORMED

11-18-82 THROUGH 11-22-82.

3.4 SAFETY

The safety requirements of UTD 131-012 shall be observed.

3.5 EQUIPMENT REQUIREMENTS

A listing of all test equipment used for measuring the recorded data during the test, model and serial numbers, and calibration due dates shall be maintained.

The IDM GSE unit will be used to simulate the S/C interface. The GSE unit, in conjunction with the HP9825A computer, is used to provide functional testing of the instrument during the EIAT. All data print-outs will be marked with the date and test temperature and maintained in the IDM data file.

The GSE Stimulator unit will be used to provide sensor stimulation during the EPET.

3.5.1 DUAL CURRENT SOURCE CALIBRATION

- a. Apply power to the current source.
- b. Monitor the + and - output jacks with a calibrated Keithley pico-ammeter.
- c. The measured currents versus current source switch settings shall agree with the following data within $\pm 5\%$ from 10^{-6} to 10^{-10} amps and within $\pm 8\%$ at 10^{-11} amps. (The dual current source will not be used as an absolute calibration of the log amps. It is used to provide relative differences between the two test currents by precisely varying the input voltage to two matched hi-meg resistors.)

VOLTAGE	RESISTANCE	DEVIATION	AMPS	
			+	-
1 V	10^{-6}	None	10^{-6}	10^{-6}
"	10^{-7}	"	10^{-7}	10^{-7}
"	10^{-8}	"	10^{-8}	10^{-8}
"	10^{-9}	"	10^{-9}	10^{-9}
"	10^{-10}	"	10^{-10}	10^{-10}
"	10^{-11}	"	10^{-11}	10^{-11}

- d. Measure the change in current at the + jack while depressing each of the "deviation" switches. The change in current shall correspond to the nominal deviation within $\pm 2\%$.
Verify *MP*

3.6 TEST DESCRIPTION

The EIAT will consist of the three separate tests described below.

a. EPET

The Electrical Performance Evaluation Test (EPET) provides a functional test of the instrument including instrument response to simulated ion currents and data multiplexing in the various modes. The EPET provides formatted print-outs of all test data.

b. Calibration Test

The Calibration Test provides for measurement of instrument parameters critical to the electrical calibration of the instrument. Parameter values obtained in the Calibration Test performed prior to conformal coating the circuit boards are compared to the EIAT values to assure that there has been no circuit degradation during the conformal coating process.

c. Vacuum Test

The Vacuum Test verifies correct instrument response to an ion beam in a simulated near-flight environment.

4.0 PREPARATION FOR TEST

4.1 EPET

- a. Place the test chassis in the test chamber. Connect cables as follows:

GSE J1 to TEST CHASSIS J1

- b. Place the stimulation unit outside the test chamber. (Stimulation unit chassis is at sensor potential and must be isolated.)

- c. Connect collector D to the (+) stimulation jack and B to the (-) jack.

- d. Connect the GSE to the HP9825A via the HP98032A interface cable. Perform EPET test per section 5.0.

4.2 CALIBRATION TEST

- a. Disconnect cables from the stimulation unit.

- b. Connect coax cables (electrometer inputs) to the Keithley Pico-amp source and the dual current source as required. (The chassis of each current source must be isolated.) Connect test leads to the repeller and suppressor grids and a coax to the reference potential connector and route them out of the chamber access port.

- c. Allow at least 30 minutes soak time at each temperature after TEMP monitor stabilizes.

- d. Perform Calibration test per section 6.0.

4.3 VACUUM TEST

- a. Place the sensor head assembly and the ion source in the vacuum chamber and fasten to the vacuum test fixture. Adjust the source vertically until the center of the source output grid is aligned with the center of the sensor input apertures.

- b. Connect the source to the source control unit.

- c. Connect the GSE unit to the test chassis.

- d. Connect the vacuum test cable through the vacuum chamber feedthrough. Verify pin to pin connections.

- e. Connect the sensor cable to the sensor under test and then to the electronics box located outside the vacuum chamber.

f. Verify that sensor potential is not shorted to the vacuum chamber housing and that the sensor is floating.

h. Perform the vacuum test per section 7.0.

5.0 EPET

5.1 Load the IDM EPET program into the HP9825 from the tape cartridge. Enter the date, test temperature and comments.

+60° 09/18 RT ✓ -20° 10/1

5.2 Record and print 64 frames of data at each temperature.

5.2.2 Verify the following on the data print-outs:

- SENPOT switches between two voltage levels of 99 volts and 0 volts.
- LLB remains at a fixed current level of 4.5×10^{-8} amps while LLA switches between 9×10^{-8} amps and 4.5×10^{-8} amps.
- IDM/RNG indicates deviation levels of 100% and 50%.
- Data multiplexing is as indicated in UTD 133-003, Figure 1.

6.0 CALIBRATION TEST

+60° ✓ RT ✓ -20° ✓
0.4 0.4

6.1 GRID POTENTIALS

Measure and record the suppressor grid potential with a DVM. Using an oscilloscope, verify that the repeller grid is at 0 volts most of the time and switches to +2 volts at the beginning of each 8-frame sequence as indicated in UTD 133-003.

	0°	+25°	+50°
SUPPGRID	-11.87	-11.86	-11.76
REPGRID LOW	0.00	0.00	+0.11
REPGRID HIGH	+1.9 ✓	+1.9 ✓	+1.9

6.2 DATA SHEET 1

a. Drive the reference potential input through a 100 Meg resistor to the levels indicated and record the resulting sensor potentials and TLM outputs.

b. Connect pico-amp source to collector A. Connect one output of the dual current source to collector C. Set dual current source to 10^{-8} amps. Set pico-amp source to the currents indicated and record LLB data. Reverse the connections to A and C and repeat the calibration procedure for LLA.

c. Record the Drift Signal zero level.

d. Record the RPA analog data levels.

6.3 DATA SHEET 2

a. Connect the dual current source + output to A and - output to C. Record drift signal offset readings for the input current setting and + deviations given on data sheet 2.

b. Reverse the + and - current source cables and repeat data sheet 2 for - deviations.

6.4 DATA VERIFICATION

The EIAT Calibration Test data shall agree within $\pm 2\%$ with the corresponding data from the preconformal coating Calibration Test for all tests with input currents greater than 10^{-5} amps.

7.0 VACUUM TEST

7.1.1 After the chamber reaches a pressure below 2×10^{-5} Torr turn on the instrument.

7.1.2 The following tests will be performed as a minimum:

a. Make records of the Log Level and DM outputs.


RT

b. Rotate the ion source and verify that the sensor responds to changes in the ion beam angle.

RT

HP 3000A	100M	#24139	CAL DUE	9/83	
KIEHLER	PIC AMP SOURCE	MODEL 261	#00755		1-1-83
HP 1703A	SCAF		#03213		1-2-83
KIEHLER	615 ELECTROMETER	#09695			1-1-83

HILAT IDM COLLECTOR CONFIGURATION



A PLUS ANGLE YIELDS A VOLTAGE ABOVE 2.5 VOLTS WHEN MAKING OFFSET (DOUBLE ANGLE) MEASUREMENTS.

HEAT TDM CALIBRATION DATA

SHEET 1

DATE 11-19-82

BY BOB

(FG-49 - STP5)

REFERENCE
POTENTIAL (VOLTS)

SENSOR POTENTIAL

SENSOR P

0° +25° +50°

0° +25°

0	<u>.003</u>	<u>.004</u>	<u>.005</u>	<u>4.48</u>	<u>4.48</u>
2	<u>2.00</u>	<u>2.00</u>	<u>+2.00</u>	<u>3.98</u>	<u>3.98</u>
4	<u>4.00</u>	<u>4.00</u>	<u>+4.00</u>	<u>3.50</u>	<u>3.50</u>
6	<u>6.00</u>	<u>6.00</u>	<u>+6.00</u>	<u>3.00</u>	<u>3.00</u>
8	<u>8.00</u>	<u>8.00</u>	<u>+8.01</u>	<u>2.50</u>	<u>2.50</u>
10	<u>10.00</u>	<u>10.00</u>	<u>+10.01</u>	<u>2.00</u>	<u>2.00</u>
12	<u>12.00</u>	<u>12.00</u>	<u>+12.01</u>	<u>1.50</u>	<u>1.50</u>
14	<u>14.00</u>	<u>14.00</u>	<u>+14.00</u>	<u>1.00</u>	<u>1.00</u>

	0°	+25°	+50°
10 ⁻⁶	<u>4.56</u>	<u>4.50</u>	<u>4.56</u>
10 ⁻⁷	<u>3.54</u>	<u>3.56</u>	<u>3.56</u>
10 ⁻⁸	<u>2.54</u>	<u>2.54</u>	<u>2.56</u>
10 ⁻⁹	<u>1.52</u>	<u>1.48</u>	<u>1.54</u>
10 ⁻¹⁰	<u>0.52</u>	<u>0.52</u>	<u>0.52</u>

	0°	+25°	+50°
DRIFT SIGNAL ZERO	<u>2.50</u>	<u>2.50</u>	<u>2.52</u>

	0°	+25°	+50°
FILTERS	<u>0.98</u>	<u>0.98</u>	<u>0.98</u>
TEMP	<u>1.94</u>	<u>1.94</u>	<u>1.94</u>
RPA	<u>2.94</u>	<u>2.94</u>	<u>2.94</u>
SWP	<u>3.96</u>	<u>3.96</u>	<u>3.94</u>
ELECT	<u>4.98</u>	<u>4.97</u>	<u>4.97</u>

RELAT IDN CALIBRATION DATA

DATE 11-22-82

SHEET 2

BY TES

DEVIATION

POLARITY +

I_{in}	% DEVIATION	0°	125°	150°	RNG
10^{-6}	100	3.38	3.38	3.38	4 ✓
"	70	3.18	3.20	3.18	4 ✓
"	50	4.55	4.55	4.55	3 ✓
"	30	3.83	3.83	3.83	3 ✓
"	20	3.42	3.42	3.42	3 ✓
"	10	4.40	4.42	4.42	2 ✓
"	0	2.40	2.44	2.46	1 ✓
10^{-7}	100	3.38	3.38	3.38	4 ✓
"	70	3.18	3.20	3.18	4 ✓
"	50	4.56	4.56	4.56	3 ✓
"	30	3.84	3.84	3.84	3 ✓
"	20	3.44	3.44	3.44	3 ✓
"	10	4.44	4.44	4.44	2 ✓
"	0	2.52	2.54	2.57	1 ✓
10^{-8}	100	3.38	3.38	3.38	4 ✓
"	70	3.18	3.20	3.18	4 ✓
"	50	4.54	4.54	4.54	3 ✓
"	30	3.82	3.82	3.82	3 ✓
"	20	3.42	3.42	3.42	3 ✓
"	10	4.38	4.40	4.39	2 ✓
"	0	2.30	2.37	2.42	1 ✓
10^{-9}	100	3.38	3.38	3.38	4 ✓
"	70	3.18	3.20	3.18	4 ✓
"	50	4.56	4.56	4.54	3 ✓
"	30	3.84	3.84	3.82	3 ✓
"	20	3.44	3.44	3.42	3 ✓
"	10	4.46	4.43	4.42	2 ✓
"	0	2.64 *	2.60	2.54	1 ✓
10^{-10}	100	3.38	3.38	3.33	4 ✓
"	70	3.18	3.20	3.14	4 ✓
"	50	4.55	4.56	4.40	3 ✓
"	30	3.85	3.84	3.74	3 ✓
"	20	3.44	3.45	3.36	3 ✓
"	10	4.34 *	4.52	4.42	2 ✓ *
"	0	*	*	*	1 *

HILAT EDM CALIBRATION DATA

DATE 11-27-82

SHEET 2

BY JLD

DEVIATION —
POLARITY —

I_{10}	% DEVIATION	0	425'	450'	RNG
10^{-6}	100	1.66	1.65	1.64	4 ✓✓
"	70	1.86	1.85	1.85	4 ✓✓
"	50	0.48	0.48	0.47	3 ✓✓
"	30	1.20	1.20	1.20	3 ✓✓
"	20	1.60	1.60	1.60	3 ✓✓
"	10	0.60	0.60	0.60	2 ✓✓
"	0	2.50	2.52	2.56	1 ✓✓
10^{-7}	100	1.64	1.65	1.64	4 ✓✓
"	70	1.86	1.85	1.85	4 ✓✓
"	50	0.48	0.47	0.46	3 ✓✓
"	30	1.20	1.20	1.19	3 ✓✓
"	20	1.60	1.60	1.60	3 ✓✓
"	10	0.58	0.58	0.58	2 ✓✓
"	0	2.46	2.48	2.51	1 ✓✓
10^{-8}	100	1.66	1.65	1.65	4 ✓✓
"	70	1.86	1.86	1.85	4 ✓✓
"	50	0.50	0.48	0.48	3 ✓✓
"	30	1.22	1.20	1.20	3 ✓✓
"	20	1.62	1.61	1.60	3 ✓✓
"	10	0.65	0.64	0.64	2 ✓✓
"	0	2.70	2.66	2.68	1 ✓✓
10^{-9}	100	1.66	1.65	1.66	4 ✓✓
"	70	1.86	1.86	1.86	4 ✓✓
"	50	0.48	0.48	0.49	3 ✓✓
"	30	1.20	1.20	1.20	3 ✓✓
"	20	1.60	1.60	1.60	3 ✓✓
"	10	0.59	0.61	0.63	2 ✓✓
"	0	2.49 *	2.57	2.64	1 ✓✓
10^{-10}	100	1.67	1.66	1.71	4 ✓✓
"	70	1.88	1.87	1.90	4 ✓✓
"	50	0.56 *	0.52	0.64	3 ✓✓
"	30	1.26 *	1.28	1.32	3 ✓✓
"	20	1.67 *	1.66	1.70	3 ✓✓
"	10	0.88 *	0.80 *	0.88	2 ✓✓
"	0	X	*	X	1 NOT

SECTION 4

DIAGRAMS, SCHEMATICS, DRAWINGS

IDM DRAWING LIST

- TOP ASSEMBLIES -

133-500	SENSOR HEAD ASSY
PL133-500	SENSOR HEAD ASSY
133-506	SENSOR INSTALLATION
133-565	SENPO T R.F. BYPASS ASSY (PL INCLUDED ON ASSY DWG)
D1293	INTERCONNECTION DIAGRAM (AFGL DRAWING)

- SUBASSEMBLIES -

DNA-1	MUX/DRIFT SIGNAL AMPLIFIER	
	BOARD ASSY	133-570
	PART LIST	PL133-570
	SCHEMATIC	133-572
	WIRE LIST	133-573

DNA-2	DATA STORE	
	BOARD ASSY	133-575
	PART LIST	PL133-575
	SCHEMATIC	133-577
	WIRE LIST	133-578

DNA-3	ELECTROMETER/AMPLIFIER	
	BOARD ASSY	133-580
	PART LIST	PL133-580
	SCHEMATIC	133-582
	WIRE LIST	133-583

END

FILMED

5-84

DTIC



